

CORRIGENDUM

NO- 403 dt. 15-01-19

This is for information of all the bidders that following amendments/Prebid Response/Corrigendum are being made in tender documents vide Tender No.9441/29.12.2018 of CDM&PHO, Jajpur for selection of bidders for the supply, installation of "ULTRASOUND MACHINE WITH COLOR DOPPLER".

The bidders are advised to take into account the following amendment/ Prebid Response/Corrigendum before submission of their bids against this tender. If any bidder has already submitted his/her bid, then he/she should resubmit his/her bid taking into account following amendments/Prebid Response/Corrigendum. The last date & time of submission of bid is being extended for 7 days i.e. 31.01.2019 till 4 PM and date & time of opening of bid will be on 01.02.19 at 3 PM. The other term and conditions will remain same.

Technical specification point	Technical specification	Technical specification amendment
2(15)	System should have 19" HD (1280*1024) display with tilt and swivel facility along with alphanumeric keyboard and TFT screen operator console with illuminating keys and status function	System should have 19" or more HD LED Display

Baish
15/1/19
Chief District Medical & Public Health Officer,
Jajpur