



OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH
OFFICER: JAJPUR



Tender Call No. / 829_/CDM&PHO, Jajpur dated the 08.03.2019

TENDER CALL NOTICE

Tender Call for	: Supply of Equipments , Furniture & BCL under KMC, Jajpur for the year 2018-2019
Date of Tender call	: 08.03.2019
Last Date & Time of submission of Tender	: 25.03.2019/ 12.00 noon
Date & Time of opening of Tender	: 26.03.2019/ 3.00 PM

Sealed tenders are invited from different bidders having valid GST Registration certificate for the **"Supply of Equipments , Furniture & BCL under KMC, Jajpur for the year 2018-2019"** as per the specification given below by the Chief District Medical & Public Health Officer, Jajpur. Interested bidders may submit their quotation to the under signed as per the scheduled date & time mentioned above.


Terms & conditions:

1. The undersigned shall have the right for rejecting all or any of the quotation without assigning any reason thereof.
2. Any tender received after the due date & time will be rejected. The tenders will be received through Regd. Post / Speed Post only.
3. Tender document fee of Rs.500/- in shape of Demand Draft/ Original Money Receipt which is not refundable.
4. An EMD of Rs 5000/- (Rupees Five Thousand) only will be paid in the shape of demand Draft only in favour of CDM&PHO, Jajpur from any Nationalized / Scheduled Bank payable at Jajpur Town, which will be refundable. The EMD of the unsuccessful tenderers will be returned back without interest and EMD of successful tenderer will be returned after successfully supply of purchase order.
5. The Firm should provide the List of items quoted as per the format provided (**Annexure II**).
6. Manufacturing unit who has been blacklisted either by the Tender inviting authority or by any state Govt. or Central Govt. organization is not eligible to participate in the tender for that item during the period of blacklisting. (**Annexure - III**).
7. Annual Average Turnover of Rs. 1 Crore or more in last 3 financial years of bidders. (**Annexure IV**).
8. The Original Tender Copy with Conditions and the schedules signed by the tenderer at the bottom of each page with his official seal duly affixed.
9. The rates should be quoted including F.O.R destination and excluding taxes.
10. The bidder should submit attested copy of valid GST Registration certificate & PAN along with the quotation.
11. The bidder should provide the copy of GSTR 3B along with the quotation.

12. Performance Statement (Annexure - V) during the last two years towards proof of supply of quoted items to any Govt. organization / Corporate Hospitals / PSU Hospitals / UN Agencies. The copy of Purchase orders and certificate from the user should be furnished in support of the information provided in the performance statement
13. The bidder should submit the relevant documents as per the items technical requirement provided in the technical Specification Para.
14. The firm should submit the IT return copy of last three years.
15. Two stage bidding systems will be done. (Technical bid & Price bid)
16. The bidder(s) are to submit their tenders in separate sealed covered envelopes for technical bid and commercial bid by super scribing Cover "A" (Technical Bid) & Cover "B" (Price Bid) and both the sealed covers should be put into a third outer Cover, which should be super scribed as "**Tender for Supply of Equipments, Furniture & BCL under KMC, Jajpur for the year 2018-2019.**"
17. The financial bid of only those bidders should be opened who qualify in their technical bid.
18. The price bid is to be quoted as per **Annexure - VI** attached.
19. The rate contract once approved should remain valid for one year from the date of approval.

List of Items:


(As per Annexure - I attached)

for 
Chief District Medical & P.H Officer, Jajpur

Memo No_830_/CDM& PHO, Jajpur dated the 08.03.2019

Copy submitted to DIO, NIC, Jajpur for favour of information and necessary action. He is requested to upload the quotation on the district website www.jajpur.nic.in for wide publication.

Caption: "Supply of Equipments, Furniture & BCL under KMC, Jajpur for the year 2018-2019"

for 
Chief District Medical & P.H Officer, Jajpur

List of Items (with technical specification)

1. Semi Recliner Bed
2. Recline Chair
3. Bag & Mask (Size 0 & 1)
4. Digital Weighing Machine
5. Digital Room Thermometer
6. Training Manikins
7. Mattress
8. Disposable Draw Sheet
9. Disposable Gown
10. Pillow
11. Bedsheets
12. Adult Blankets
13. Baby Blankets

for

02/03/19

Chief District Medical & P.H Officer, Jajpur

Equipment Specification for KMC Unit

1. Semi Recliner Bed

Quality standards:

- Manufacturer should be ISO 9001 certified.
- Should have ISO14001- for Environment friendly features.
- Should have ISO18001for occupational health safety management
- Product must be CE/BIS/BIFMA certified.
- Should furnish stainless steel grade certificate from Govt/Govt. approved testing laboratory.
- Manufacture should produce test certificate from Govt/Govt. approved laboratory for test procedure like impact test, bend test, salt spray chamber test, epoxy powder coating & phosphate coating for quoted item
- Semi fowlers' bed made of Rectangular M.S.Tube with two sections Top.
- The top should be made of CRCA perforated sheet.
- Back rest section has to be manoeuvred by screw handle at foot end up to an Angle of 60degree
- S.S laminated Head and Foot boards.
- Bed Mounted On 100mm Dia caster.
- Approximate Diamention:200Lx90wx 60H
- Finish should be Pre treated & Epoxy powder coated

2. Recline Chair

Technical Specification:

- It should be a folding gravity recliner chair.
- Smooth & easy reclining mechanism provides almost sleeping body posture
- Unique and secured locking mechanism locks reclining position at any stage to provide desired body posture
- Strong and stable design gives long lasting service
- Easy to fold and unfold
- UNIQUE, STRONG CAPSULE TUBE CROSS FRAME ENABLES TO TAKE LOAD UP TO 130 Kgs
- Frame : Powder coated round tubular Mild Steel
- Frame Fabrics: High Quality Taxylene/ equivalent.
- Gross Weight : 8 Kgs.(Approximate)
- Dimensions (Approximate): Seat :160 x 50 Cms Seat Height : 40 Cms
- Manufacturer should be ISO certified.

3. Bag & Mask (size 0 & 1)

Clinical purpose:

- To provide or assist ventilation in a patient who is apnoeic or exhibits inadequate respiration through manual pulmonar-driven pressure cycle functions.
- **Used by clinical department/ward**-It is used in ambulances, intensive care units (ICU), during internal patient transfer, accident and emergency (A&E), and mass casualty incidents (MCI).

Overview of functional requirements:

Technical Specification:

- Hand held Manual resuscitator with transparent face-mask.
- Child models (750ml, 500ml and 260ml bag capacity);
- Standard 15/22 mm Swivel connector allows connections to all common masks Endo-tracheal Tubes;
- Provision to give supplemented oxygen-by-oxygen reservoir providing 100% oxygen;
- Non-re breathing valve enabling the patient to inspire oxygen from the reservoir bag.
- Should be single hand opera table
- Should be easy to disassemble for cleaning and disinfection
- Should have pressure release valve at 40cm H₂O
- Should have silicone oxygen tube 2m length.
- It should be upto 40 times autoclavable including bag and washers.
- The bag should be made of soft silicone material.
- Self Inflating Resuscitator bag should be of medical grade silicone rubber.
- The reservoir should be a PVC bag of 600ml capacity for 260ml & 500ml bag capacity and 1000ml for 750ml bag capacity.
- **Accessories:** Silicon bellow, Non Rebreathing Valve, 2 meter oxygen tube, Guedel Airway, Oxygen Reservoir bag and Neonatal Mask of 2 sizes viz 0 and 1

Environmental and Departmental Considerations:

Atmosphere / Ambiance (air conditioning, humidity, dust ...)

- **Operating condition:** –Capable of operating continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances.
- An ambient air velocity is less than 0.3 m/s.

User's care, Cleaning, Disinfection & Sterility issues

- Complete unit to be easily washable and sterilizable using both alcohol and chlorine agents.

Standards and Safety:

- Certificates (pre-market, sanitary, ..); Performance and safety standards (specific to the device type);
- Local and/or international-ISO 13485; Manufacturer / supplier should have ISO certificate for quality standard.
- Should be USFDA / CE(notified) approved product or BIS certified
- Should meet ISO 10651-4 standard requirement

Recommendations for maintenance-NA

Notes:

- Service Support Contact details (Hierarchy Wise; including a toll free/landline number)-NA
- Recommendations or warnings-NA

4. Digital Weighing Machine

- Clinical purpose -To measure body mass of the neonate
- Used by clinical department/ward-NICU/SNCU/KMC

Technical specification:

- Table top, light and portable
- Built in rechargeable battery
- Easy to clean baby tray (acrylic)
- Zero weight adjustment facility
- Quick, clear digital read outs
- Measurement does not change with position of baby on the pan;
- Provision to measure the height of the baby in its laying position
- Accuracy: 5g, resolution: 1g, limit: 10gm ~ 15kg
- Settings-Auto setting to 0.00 once a the machine is switched on or when no external weight has been put on
- User's interface-LCD display
- Software and/or standard of communication(where ever required)-in built

Physical Characteristics:

- Dimensions (metric)- Base: 300mm x 265mm x 85mm, Pan: 510mm x 300mm x 85mm
- Mobility, portability-portable

Energy Source:

- Power Requirements-230 V AC,
- Battery operated-6V, one hour backup

Environmental and Departmental Considerations:

- **Atmosphere / Ambiance (air conditioning, humidity, dust ...)**-Operating condition: Capable of operating continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances. an ambient air velocity less than 0.3 m/s.
- **User's care, Cleaning, Disinfection & Sterility Issues**-Complete unit to be easily washed and disinfected using both alcohol and chlorine agents.

Standards and Safety:

- **Certificates (pre-market, sanitary...); Performance and safety standards (specific to the device type); Local and/or international**-The Scale should be as per BIS specifications. The scale should have ISI mark ie IS: 2489 Or CE/FDA certified. Should have model approval from Legal Metrology Dept., Govt. of India.
- The manufacturer shall have the valid license and should have model approval by the legal metrological Deptt. and the weighing scale must be stamped by the by legal metrological Deptt. In case of distributor, the bidder should have valid distributor and repair license from legal metrological Deptt., Govt. of Odisha.

5. Digital Room Thermometer

Technical Specification:

- To measure the temperature and humidity of the indoor and outdoor
- The temperature measurement range:
- The measurement unit of the temperature $^{\circ}\text{C}/^{\circ}\text{F}$ changeable
- For Indoor: $-10^{\circ}\text{C} \sim 50^{\circ}\text{C}$ ($+14^{\circ}\text{F} \sim 122^{\circ}\text{F}$)
- For Outdoor: $-50^{\circ}\text{C} \sim 70^{\circ}\text{C}$ ($-58^{\circ}\text{F} \sim 158^{\circ}\text{F}$)
- The Humidity Range: indoor 20% ~ 99% RH
- The Storage condition of the machine : $-20\text{C} \sim 60\text{C}$, 20~80%RH
- The unit shall work on battery.
- Easy user setting and should give reading in less time.
- All measurement should display in a clear LCD screen
- The manufacturer should be ISO certified.
- The unit should supplied with battery.

6. Training manikins

FBNC-Normal New Born Baby Simulation Model for KMC:

Technical Specification:

- The material of mannequin should be of polyvinyl or silicon rubber, free from any hazardous material.
- The texture of the mannequin should be as close to the feel of the baby/adult skin as relevant

- New born baby mannequin should weigh close to the normal newborn.
- Should have actual size showing external development and growth.
- Should be close to normal skin colour, texture and bony feel.
- Should have moving head, flexible upper and lower limbs.
- Should have KMC clothes compatible with size of the mannequins.
- Standards:
- BS EN ISO/ IEC 17050-1:2010
- Conformity assessment .Suppliers declaration of conformity.
- EMC Directive :2004/108/EC

7. Mattress:

- Two section mattresses with 3" thick PU foam of 40 density covered with PVC Rexene
- Approximate Dimension: 190cm Lx 80cm W
- Manufacture should be ISO Certified

8. **DISPOSABLE DRAW SHEET**

- It should be made up of SMS cloth.
- It should be good quality interlining fabric with 40 gsm capacity.
- It should be waterproof.
- It should be provided in double packing.
- The size of disposable draw sheet should be of 120 cm x 100 cm in size.
- The fabric composites should be strong durable, anti-microbial, anti-static, flame resistant, stain, odor, & fluid resistant.
- The material should be of high grade, soft texture, top quality and skin friendly.
- Product should be CE/BIS approved.
- Test Certificate for 40GSM should be provided from a certified laboratory.

9. **DISPOSABLE PATIENT GOWN**

- The Disposable Gown should be made up of non-woven cloth
- It should be of 40 GSM good quality interlining fabrication non woven.
- It should be minimum of 1.30 mtr long.
- The disposable gown should be provided with double packing.
- It should be provided with sterile pack.
- Sterilisation process – ETO/Gamma irradiation .
- Product should be CE/BIS approved.
- Test Certificate for 40GSM should be provided from a certified laboratory.

(To be submitted in Cover A -Technical Bid)

LIST OF ITEM(S) QUOTED

Sl.	Name of Item (s)	Specification	Name of Manufacturer	Remarks

Signature of the Tenderer:

Date:

Seal

(To be submitted in *Cover A - Technical Bid*)

DECLARATION FORM

I / Wehaving My / our
.....office at.....do
declare that I / We have carefully read all the terms & conditions of tender of the _____,
Odisha for the **“Supply of Equipments , Furnitures & BCL items under KMC, Jajpur for the
year 2018-2019”**. The approved rate will remain valid for a period of one year from the date of
approval. I will abide with **all the terms & conditions** set forth in the **Tender Reference no.**

I/We do hereby declare I/We have not been de-recognised / black listed by any State
Govt. / Union Territory / Govt. of India / Govt. Organization / Govt. Health Institutions for
supply of Not of Standard Quality (NSQ) items / non-supply.

I/We agree that the Tender Inviting Authority can forfeit the Earnest Money Deposit
and or Performance Security Deposit and blacklist me/us for a period of 3 years if, any
information furnished by us proved to be false at the time of inspection / verification and not
complying with the Tender terms & conditions.

I / We do hereby declare that I /
we will supply the _____ as per the terms, conditions & specifications of the tender
document. .

Signature of the bidder:

Seal

Date :

Name & Address of the Bidder:

Affidavit before Executive Magistrate / Notary Public.

(To be furnished in the **letter head** of the Auditor/ Chartered Accountant)

ANNUAL TURN OVER STATEMENT

The Annual Turnover for the last three financial years of M/s _____ who is a Manufacturer /Distributor/Importer (*Pl. tick whichever is applicable*) are given below and certified that the statement is true and correct.

Sl.No.	Year	Turnover in (Rs.)
1.	2015-2016	-
2.	2016-2017	-
3.	2017-2018	-

Average Annual Turnover
(for the above three years) in **(Rs.)**

Date:
Place:

Seal

Signature of Auditor/
Chartered Accountant
(Name in Capital)

Membership No.-

Registration No. of Bidder

Note:

- a) To be issued in the **letter head** of the Auditor/Chartered Accountant mentioning the Membership no.
- b) **Separate certificates** should be furnished for **different manufacturer/importer** in case the bidder (authorized distributor) is quoting products of **different manufacturers/importers**. The authorized distributor has also to furnish his turnover statement in the above format.

(To be submitted in **Cover A - Technical Bid**)
PROFORMA FOR PERFORMANCE STATEMENT
 (For the period of last **two years**)

Tender Reference No. : _____
 Name of Tenderer : _____
 Name of Manufacturer : _____

Sl.	Order placed by (Address of purchaser) (attach documentary proof)*	Order no. & Date	Item Name	Manufacturer	Qty	Value of Contract (Rs.)
1						
2						
..						

Signature and seal of the Tenderer

- * The documentary proof will be copies of the purchase order (during the last 2 years).

(To be submitted in *Cover B - Price Bid*)**LIST OF ITEMS**

Sl No	Name of the item	Specification (As per Annexure - I)	Mfd. Name	Cost (in Rs) (including F.O.R destination)	GST + OT (if any)	Total Cost
1	Semi Recliner Bed					
2	Recline Chair					
3	Bag & Mask (Size 0 & 1)					
4	Digital Weighing Machine					
5	Digital Room Thermometer					
6	Training Manikins					
7	Mattress					
8	Disposable Draw Sheet					
9	Disposable Gown					
10	Pillow					
11	Bedsheets					
12	Adult Blankets					
13	Baby Blankets					

for *02/03/19*
Chief District Medical & P.H Officer, Jajpur